



2018年5月10号实验室安全与环保检查情况表

楼名/房间	安全隐患	整改情况	05.10 检查图片
西四教	实验室桌面脏乱	已提醒	
西四教	钢瓶未固定	已提醒	

西四教	实验室化学药品无动态清单	已提醒	
西四教	钢瓶未标明气体,无安全保障; 实验室内卫生较差; 化学药品 无动态清单	已提醒	 

<p>西四教</p>	<p>实验室内卫生较差；吹风机用完未拔插头</p>	<p>已提醒</p>	 <p>The top photograph shows a purple and white hair dryer plugged into a power outlet on a laboratory bench. In the background, there are various chemical bottles and containers, including one labeled 'AIBN'. The bottom photograph shows a green trash bin overflowing with white paper waste, with more trash scattered on the floor nearby.</p>
<p>西四教</p>	<p>钢瓶无安全保障；实验室内卫生较差</p>	<p>已提醒</p>	 <p>The left photograph shows a tall, rusted metal gas cylinder standing upright in a laboratory. The right photograph shows a cluttered laboratory workbench with various items, including a fire extinguisher, boxes, and other equipment.</p>

<p>逸夫楼 406</p>	<p>实验室内卫生较差；无药品动态清单</p>	<p>第二次提醒</p>	 <p>The top photograph shows a laboratory bench with a computer monitor, keyboard, and various bottles and containers scattered on the floor and on the bench. The bottom photograph shows a laboratory shelf filled with numerous bottles and containers, with a large white jug and green gloves on the bench in the foreground.</p>
----------------	-------------------------	--------------	--

<p>逸夫楼</p>	<p>实验室无人没有关门；实验室 内卫生较差</p>	<p>已提醒</p>	
------------	--------------------------------	------------	---

学院楼	实验室内卫生较差，瓶子摆放过多	已提醒	
学院楼	钢瓶未固定	已提醒	

学院楼	实验室内卫生较差	已提醒	
学院附楼	化学药品无动态清单	已提醒	